Bed, Delivery (Motorized)

S.N.	Purchaser's Specifications		Bidder's Offer
	Bed, Delivery (Motorized)		
	Manufacturer		
	Brand		
	Type/Model		
	Country of Origin		
1.	**Description of Function**		
1.1	Delivery bed is used for Baby Delivery and must incorporate ideal blend of the patient's comfort and the professional needs of the delivery team.		
2.	**Operational Requirements**		
2.1	Fully motorized elcetric delivery bed with all attachments and positions.		
3.	**System Configuration**		
3.1	Delivery Bed (Motorized) with all attachments and accessories.		
4.	**Technical Specifications**		
4.1	The table shall be mobile on castors with efficient braking system or castor lifting system in order to provide sufficient stability during operation.		
4.2	Three sections top made of Stainless Steel (AISI 304) or ABS material or equivalent.		
4.3	User friendly handset for electrically adjustments of backrest, height and trendelenburg / reverse trendelenburg.		
4.4	Easily removable head and foot boards.		
4.5	Foot support for high /low chair position.		
4.6	Middle and leg section tops removable for cleaning.		
4.7	Corner buffers.		
4.8	All mild steel parts are pre-treated and powder coated finish.		
4.9	Battery back-up with built in charger with AC adapter for minimum three hours of operation.		
4.10	Bed shall have a load capacity of at least 200 kg.		
4.11	Mattress: antistatic, 3 section, washable and easy to clean with a thickness of at least 7cm.		
4.12	It should have side rails.		
5.	**Accessories, spares and consumables**		
5.1	**Accessories:** • A pair of upholstered aluminium lithotomy crutches mounted on stainless steel rods. • Stainless Steel telescopic I.V. Pole -1 set. • Stainless Steel Tray-1set. • Dust protective base cover-1pc. • Patient grip rod on both sides		
5.2	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above).		
6.	**Operating Environment**		
6.1	The system offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The		

S.N.	Purchaser's Specifications	Bidder's Offer
	conditions include Power Supply, Climate, Temperature, Humidity, etc.	
6.2	Power supply: 220 – 240 VAC, 50Hz Single Phase fitted with appropriate plug. The power cable must be at least 3 metre in length.	
7.	**Standards and Safety Requirements**	
7.1	Must submit ISO 13485:2016 **AND**	
7.2	CE or USFDA approved certificate.	
7.3	Shall meet IEC 60601-2-46 Medical Electrical Equipment - PART 2-46: Particular Requirements for the Safety of Operating Tables.	
8.	**User Training**	
8.1	Must provide user training (including how to use and maintain the equipment).	
9.	**Warranty**	
9.1	Comprehensive warranty for 2 years.	
10.	**Maintenance Service During Warranty Period**	
10.1	During warranty period supplier must ensure corrective/breakdown maintenance whenever required.	
11.	**Installation and Commissioning**	
11.1	The bidder must arrange for the equipment to be installed and commissioned by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.	
12.	**Documentation**	
12.1	User (Operating) manual in English.	
12.2	Service (Technical / Maintenance) manual in English.	
12.3	List of important spare parts and accessories with their part numbers and costing.	
12.4	Certificate of calibration and inspection from factory.	