Plaster Cutter

S.N.	Purchaser's Specifications		Bidder's Offer
	Plaster Cutter		
	Manufacturer		
	Brand		
	Type / Model		
	Country of Origin		
1	**Description of Function**		
1.1	To cut all kinds of plaster bandages.		
2	**Operational Requirements**		
2.1	Plaster Cutter with blades of different diameter.		
3	**System Configuration**		
3.1	Plaster Cutter, complete unit with all standard accessories.		
4	**Technical Specifications**		
4.1	Must have a fibre body.		
4.2	Must be able to cut fibre gauzes.		
4.3	Blades must be corrosion resistant and highly durable and autoclaveable.		
4.4	Blades must have hexagonal mounting hole.		
4.5	Must be supplied with 84mm, 74mm, and 64mm diameter blades.		
4.6	Must have a protective guard.		
4.7	Must be oscillating type.		
5	**Accessories, spares and consumables**		
5.1	**Accessories:** • Blades (84mm, 74mm, 64mm) – 06 each. • Tools for replacing the blades, brush and duster. • Carrying case to accommodate the plaster cutter and other accessories.		
5.2	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above).		
6	**Operating Environment**		
6.1	The system offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Power Supply, Climate, Temperature, Humidity, etc.		
6.2	Power supply: 220 – 240 VAC, 50Hz Single Phase fitted with appropriate plug. The power cable must be at least 3 metre in length.		
7	**Standards and Safety Requirements**		
7.1	Must submit ISO 9001 or ISO13485:2003/AC:2007 for Medical Devices **AND**		
7.2	CE or USFDA approved product certificate.		
7.3	Electrical safety conforms to standards for electrical safety IEC 60601-1 General requirement for Electrical safety of Medical Equipment.		
8	**User Training**		
8.1	Must provide user training (including how to use and maintain the equipment).		
9	**Warranty**		
9.1	Comprehensive warranty for 2 years after acceptance.		

S.N.	Purchaser's Specifications	Bidder's Offer
10	**Maintenance Service During Warranty Period**	
10.1	During warranty period supplier must ensure preventive maintenance & corrective/breakdown maintenance whenever required.	
11	**Installation and Commissioning**	
11.1	The bidder must arrange for the equipment to be installed and commissioned by certified or qualified personnel; any prerequisites for installation to be communicated to the purchaser in advance, in detail.	
12	**Documentation**	
12.1	User (Operating) manual in English.	
12.2	Service (Technical / Maintenance) manual in English.	
12.3	Certificate of calibration and inspection from factory as per manufacturer guidelines.	