Trolley, Medicine

S.N.	Purchaser's Specifications	Bidder's Offer
	Trolley, Medicine	
	Manufacturer	
	Brand	
	Type / Model	
	Country of Origin	
1	**Description of Function**	
1.1	A medicine/drug trolley for storage and delivery of medicines and drugs to patients in wards of healthcare facilities.	
2	**Operational Requirements**	
2.1	Stainless steel medicine trolley with swivel castors.	
3	**System Configuration**	
3.1	Medicine Trolley, complete unit.	
4	**Technical Specifications**	
4.1	It shall be constructed fully with 304 grade stainless steel sheet and tube or better.	
4.2	Overall size: approximately 900 H x 460 W x 760 L mm.	
4.3	Flat top of SS and at least 6 inches deep removable bucket at bottom.	
4.4	Multiple drawers (minimum 6) with telescopic channels, below the platform.	
4.5	Minimum 6 small boxes to keep drugs at eye level.	
4.6	Equipped with waste bin, needle disposable container, file cassette, atraumatic corner buffers and guard rails.	
4.7	Shall have provision for hanging one IV fluid bottle.	
4.8	Equipped with lock key system.	
4.9	Shall be mobile on 4 x 125mm (+/- 5%) diameter (approx.) robust 360 deg. anti-rust, anti-static, noiseless, swivel castors with non-marking tyres and with at least 2 diagonal castors shall have brakes.	
5	**Accessories, spares and consumables**	
5.1	**Accessories:** • All standard accessories/ consumables/ parts required for the proper operation of the above item shall be included in the offer. Bidders shall specify, in a separate Excel worksheet, the quantity and details of any items included in this offer which have not been specified in this Technical Specifications Form.	
6	**Operating Environment**	
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.	
7	**Standards and Safety Requirements**	
7.1	Must submit ISO 13485:2016 for Medical Devices **AND**	
7.2	CE or USFDA approved product certificate.	
8	**User Training**	
8.1	Not applicable.	
9	**Warranty**	
9.1	Warranty for 2 years after acceptance.	
10	**Maintenance Service During Warranty Period**	
10.1	Standard warranty conditions are applicable.	
11	**Installation and Commissioning**	
11.1	Must supply preassembled unit, ready to use.	
12	**Documentation**	
12.1	User's manual shall be supplied in English.	