Wheel Chair

S.N.	Purchaser's Specifications		Bidder's Offer
	Wheel Chair		
	Manufacturer		
	Brand		
	Type / Model		
	Country of Origin		
1	**Description of Function**		
1.1	Wheel chair is used in hospitals for means of mobility by disabled persons/or persons who have impairments that limit their ability to walk.		
2	**Operational Requirements**		
2.1	It shall be a foldable BUT shall NOT a collapsible type. The mechanism of folding & unfolding must be easy. Easy maneuverable.		
3	**System Configuration**		
3.1	Wheel chair invalid type.		
4	**Technical Specifications**		
4.1	Must be made of the highest quality materials such as Chrome polished finish or stainless steel.		
4.2	Dimensions: approx. W 68 cm × D 110 cm × H 94 cm. Seat width: approx.450mm.		
4.3	Wheels to have braking/locking mechanism and self-propelling SS hoops; two swivel castors (200mm dia. approx.) in front.		
4.4	Tire fitted with self-propelling hoops and brake arrangements.		
4.5	Tire sizes: Rear: solid Mag tyres wheels, and Front: Mag swivel casters.		
4.6	Armrests: Padded, Fixed height and detachable.		
4.7	Waterproof upholstery and easy to clean.		
4.8	Padded back rest, seat and push handle.		
4.9	Footrests: Fixed height and swing away foot plates and detachable, preferably made of Aluminium.		
4.10	Maximum Patient weight capacity: approx. 150kg.		
4.11	I.V. pod shall be provided at the right side of the back rest.		
5	**Accessories, spares and consumables**		
5.1	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above).		
6	**Operating Environment**		
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.		
7	**Standards and Safety Requirements**		
7.1	Must submit ISO 13485:2016 for Medical Devices **AND**		
7.2	CE or USFDA approved product certificate.		
8	**User Training**		
8.1	Not applicable.		
9	**Warranty**		
9.1	Comprehensive warranty for 2 years.		

S.N.	Purchaser's Specifications	Bidder's Offer
10	**Maintenance Service During Warranty Period**	
10.1	Standard warranty conditions are applicable.	
11	**Installation and Commissioning**	
11.1	Must supply preassembled unit, ready to use.	
12	**Documentation**	
12.1	User's manual shall be supplied in English.	