Laryngoscope Set, Children & Adult

S.N.	Purchaser's Specifications		Bidder's Offer
	Laryngoscope Set, Children & Adult		
	Manufacturer		
	Brand		
	Type / Model		
	Country of Origin		
1	**Description of Function**		
1.1	Laryngoscopy to assist endotracheal intubation of adults, children and infants during anaesthesia or resuscitation.		
2	**Operational Requirements**		
2.1	Battery-powered laryngoscope unit		
3	**System Configuration**		
3.1	Laryngoscope set for children & adults, complete unit.		
4	**Technical Specifications**		
4.1	It shall have a large hollow, cylindrical, non-slippery handle.		
4.2	Shall be made up of Surgical Grade Stainless Steel (AISI 304/316 or equivalent).		
4.3	It shall have provision to insert battery / batteries		
4.4	It shall have provision for fitting various sizes and types of depressors/blade.		
4.5	Light to be activated when the depressor/blade is engaged.		
4.6	Shall come with a set of four stainless steel depressor/blade, with LED bulb: **Macintosh type:** • Curved Nr 2, length approx. 110 mm • Curved Nr 3, length approx. 135 mm • Curved Nr 4, length approx. 155 mm **Miller type:** • Straight Nr 1, length approx. 100 mm		
4.7	It shall be autoclavable /sterilizable.		
5	**Accessories, spares and consumables**		
5.1	**Accessories:** • Spare bulbs: 3 nos. • Spare batteries: 2 nos. • Durable protective box or case: 1 no.		
5.2	All standard accessories, consumables and parts required to operate the equipment, including all standard tools and cleaning and lubrication materials, to be included in the offer. Bidders must specify the quantity of every item included in their offer (including items not specified above).		
6	**Operating Environment**		
6.1	The product offered shall be designed to be stored and to operate normally under the conditions of the purchaser's country. The conditions include Climate, Temperature, Humidity, etc.		
6.2	It shall be battery operated system.		
7	**Standards and Safety Requirements**		
7.1	Must submit ISO13485:2016 for Medical Devices **AND**		
7.2	CE or USFDA approved certificate.		
8	**User Training**		

S.N.	Purchaser's Specifications	Bidder's Offer
8.1	Not applicable.	
9	**Warranty**	
9.1	Comprehensive warranty for 2years after acceptance.	
10	**Maintenance Service During Warranty Period**	
10.1	Standard warranty conditions are applicable.	
11	**Installation and Commissioning**	
11.1	Must supply preassembled unit, ready to use.	
12	**Documentation**	
12.1	User (Operating) manual in English.	
12.2	Letter of commitment from principal manufacturing company regarding the availability of spare parts and accessories for next five years after the completion of warranty period.	