



Government of Sudur Pashchim Province
Ministry of Social Development
Province Public Health Laboratory
Dhangadhi, Kailali Nepal

Amendment in Bid Document

Invitation to electronic bids for PPHLSP/G/NCB-01/2081-082 Procurement Supply and Delivery of Ultra-Modern Microbiology and Infectious Disease Lab setup Kartik 6th in the National News Paper, Province Public Health Laboratory, Dhangadhi has amended on bids here by: -

A. Technical Specification

For Fully Automated MalDI-TOF System

SN	Existing as per bid Document (Current)	Amended (Applicable after Amendment)
4.4	Should have oil free vacuum pumps for easy maintenance.	Should have oil based or oil free vacuum pumps.
12	Documentation	
12.3	The bidder must separately quote the price of all reagent and consumables valid for at least 2 years.	Added

For Fully Automated Microbes Identification, Antibiotics & Antifungal Drug

Sensitivity System

SN	Existing as per bid Document (Current)	Amended (Applicable after Amendment)
12	Documentation	
12.3	The bidder must separately quote the price of all reagent and consumables valid for at least 2 years.	Added

For Water bath

SN	Existing as per bid Document (Current)	Amended (Applicable after Amendment)
4.6	Must have a stirrer for circulation.	Must have a stirrer for circulation or the system must have water pump for uniform distribution of heat.

नि. निर्देशक



For Steam Autoclave

SN	Existing as per bid Document (Current)	Amended (Applicable after Amendment)
3.1	Steam autoclave, capacity approx. 100 liters or more, complete with accessories	Steam autoclave, capacity approx. 90 liters or more, complete with accessories

For Dry bath

SN	Existing as per bid Document (Current)	Amended (Applicable after Amendment)
4.2	Temperature Range: Amb +5° to 130°C	Temperature Range: Amb +5° to 80°C or more
4.6	Time Range: 0-99hrs 59 mins	Time Range: 0-99 mins or more
4.5	Block Capacity: 4	Block Capacity: 4 or more

For Digital pH meter

SN	Existing as per bid Document (Current)	Amended (Applicable after Amendment)
4.2	Mile volt range must be 0-1999+/- mv.	Removed
4.6	Temperature compensation 0-100 °C.	Temperature compensation 10-50 °C or more.
4.7	Display 4-digit LED with automatic polarity.	Display 3-digit LED or more with automatic polarity.

For Hot Plate with Magnetic Stirrer

SN	Existing as per bid Document (Current)	Amended (Applicable after Amendment)
4.3	Fitted with 3 position thermostats.	Fitted with 3 or more position thermostats or digital or better.
4.7	Heating: 300 to 340 °C.	Heating: 0 to 100 °C or more.


 नि. निर्देशक



For Hot Air Oven (Small)

SN	Existing as per bid Document (Current)	Amended (Applicable after Amendment)
4.8	Temperature Range: Ambient to 250 °C.	Temperature Range: Ambient to 200 °C or more.

B. Delivery and Completion Schedule

SN	Existing as per bid Document (Current)	Amended (Applicable after Amendment)
	Within 45 Days after signing of Agreement	Within 90 Days after signing of Agreement

Acting Director

नि. निर्देशक